ResMed

Fitting Guide

Prior to fitting your mask, remove the orange spring frame support from the rear of your mask and retain for future use.



• With both lower headgear clips released, position the cushion at an angle on the nasal bridge.



 Keeping the upper position of the mask in place, swing the lower half of the mask down against the face.



 Hold the mask firmly against the face and pull the headgear over the head.



• Bringing the lower straps below the ears, clip the headgear into the lower part of the spring frame.



• Unfasten the Velcro®, pull the upper headgear straps evenly until they are comfortable and firm, and reattach the Velcro.



• Repeat step 5 with the lower headgear straps.



- Connect one end of the air tubing to the swivel and the other end to the airflow device.
- Turn on the airflow device.

Tip One



• Ensure the headgear is correctly positioned on the head.







• Ensure the mask sits mid-bridge and does not encroach on the eyes.



www.quattrofx.com



Manufacturer: ResMed Ltd 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia. Distributed by: ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA; ResMed (UK) Ltd 96 Milton Park Abingdon Oxfordshire OX14 4RY UK See www.resmed.com for other ResMed locations worldwide. Quattro is a trademark of ResMed Ltd and Velcro is a trademark of Velcro Industries B.V. © 2010 ResMed Ltd. Specifications may change without notice. 1013934/2 2010-08